

# TEDS MODULE CONTENTS

SERIAL 1017181A  
DATE Feb-13-2018

Device Type	DS2433	(DS1973)
Manufacture ID	59	Interface Inc.
Model Number	1210	
Version Number	1	
Version Letter		
Serial Number	1017181	
Template ID	33	
Physical Measurand	4 (N)	
Minimum Physical Value	-12500	
Maximum Physical Value	12500	
Precision Selector	2 (32 Bits)	
Minimum Output	-0.00215142 V/V	
Maximum Output	0.00215291 V/V	
Bridge Type	2 (Full)	
Bridge Impedance	351.10 Ohms	
Response Time	0.000001 Seconds	
Excitation (Nominal)	10.0 Volts	
Excitation (Minimum)	5.0 Volts	
Excitation (Maximum)	15.0 Volts	
Calibration Date	Feb-13-2018	
Calibration Initials	RV	
Calibration Period	365	
Measurement Location ID	1	